



UNIVERSITY OF
GEORGIA

Center for Applied Isotope Studies
120 Riverbend Road
Athens, Georgia 30602
TEL 706-542-1395 | FAX 706-542-6106
biobase@uga.edu
www.cais.uga.edu

Certificate of Analysis

March 8, 2022

François Loin
POLLET S.A.
Rue de la Grande Couture 20
B-7501 Tournai (Orcq)
Belgium

Listed below are the isolated results for the ASTM method D6866-20 Radiocarbon (^{14}C) determination with the stable carbon isotope ratio ($\delta^{13}\text{C}$) analyses and their correction for the following sample received by our laboratory on 2/11/2022 and completed on 3/7/2022.

Sample ID/USDA#	^{14}C (Meas.)		$\delta^{13}\text{C}$	^{14}C (Corr.)	% Biobase	
	(pMC)	SD	(‰ VPDB)	(pMC)	Carbon	SD
CAPS Indoors, USDA# 10180/ 220101	40.14	0.16	-29.04	40.47	40	1

Percent Biobased Carbon is determined from the measured ^{14}C in percent Modern Carbon (pMC) and corrected for isotopic fractionation based on measured $\delta^{13}\text{C}$ value (‰ V-PDB). The corrected ^{14}C activity in pMC is then divided by the 2018 reference ^{14}C activity of 100.0 pMC, which represents the equivalence to the 1950 ^{14}C reference activity of 13.56 dpm/gC corrected for bomb-produced ^{14}C , and finally multiplied times 100. The % Biobase Carbon and Standard Deviation (SD) are rounded to the nearest integer. Measured ^{14}C is normalized using NIST Standard Reference Material 4990C Oxalic acid.

Authorized by,

Michael C Marshall, PhD
Assistant Research Scientist & Quality Manager
C.A.I.S. Inv. No: [NPI220815]
Certificate#: [POLLET_1-61246I_1159]